

*If you are using a printed copy of this procedure, and not the on-screen version, then you **MUST** make sure the dates at the bottom of the printed copy and the on-screen version match. The on-screen version of the Collider-Accelerator Department Procedure is the Official Version. Hard copies of all signed, official, C-A Operating Procedures are kept on file in the C-A ESHQ Training Office, Bldg. 911A*

C-A OPERATIONS PROCEDURES MANUAL

ATTACHMENT

4.61.a NMC Recheck Sheet

Text Page 2

C-A OPM Procedures in which this Attachment is used.		
4.61		

Hand Processed Changes

<u>HPC No.</u>	<u>Date</u>	<u>Page Nos.</u>	<u>Initials</u>
_____	_____	_____	_____
_____	_____	_____	_____
_____	_____	_____	_____

Approved: \_\_\_\_\_ *Signature on File* \_\_\_\_\_  
 Collider-Accelerator Department Chairman Date

D. Gassner

## NMC RECHECK DATA SHEET

Upon commissioning and at monthly intervals during operations of the applicable beam line, NMC monitors will have the following data recorded in a special log:

1. Date \_\_\_\_\_
2. By \_\_\_\_\_
3. S/N of NMC unit \_\_\_\_\_
4. S/N of paddle \_\_\_\_\_
5. Upscale Check voltage at input to integrator card \_\_\_\_\_
6. Voltage measured using a Fluke model 23 DVM (from TB1-5)

LAMP VOLTS TB1-4	METER READING	CONTROL VOLTS card pin M	HIGH VOLTS TB1-3
0V			
2V			
4V			
6V			

FINAL SETUP

\_\_\_\_\_